


# CORONET

## Hydroxyapatite Orbital Implants Specification Sheet



### Product Overview

All CORONET® Hydroxyapatite Orbital Implants have a 100% interconnected micro pore structure, which provides an ideal surface area for absorption of biomolecules, essential for cellular attachment and infiltration.

- Ultra-high porosity (80%>)
- Constructed with 100% medical grade synthetic Hydroxyapatite (HA)
- Recommended for PEG procedures for oculoplastic prosthesis.
- Supplied sterile, single use only, one per pack, 5 year declared shelf life

### Orbital Implant Options

Size (mm)	54-300 100% medical grade synthetic hydroxyapatite (HA)
14.00	54-300-14.00
16.00	54-300-16.00
17.00	54-300-17.00
18.00	54-300-18.00
19.00	54-300-19.00
20.00	54-300-20.00
21.00	54-300-21.00
22.00	54-300-22.00

# CORONET

## Hydroxyapatite Orbital Implants Specification Sheet

### Material Specification

#### Product Component

#### Specification

Implant	100% medical grade synthetic hydroxyapatite (HA)
Pouch	Inner Pouch, Outer Pouch - poly/tyvek peel pouch
Shelf Carton	500 micron Printed White Boxboard

### Intended Use

The Orbital Implant is intended for use as a primary orbital implant at the time of enucleation or evisceration procedures, or as a secondary/exchange implant for later socket reconstructive surgery. It is not intended for any other purpose.

### Instructions for Use

The instructions for use and suggested surgical technique are supplied in the form of a multi-lingual leaflet with instructions given in diagram form where appropriate. The international symbols used on the packaging are explained in each language. A leaflet is supplied with each product

### Sterilisation

Products are sterilised by Gamma irradiation from a Cobalt 60 source in accordance with a validated 25 kGy cycle. Sterilisation is carried out in accordance with the requirements of ISO 11137-1 (current) and 11137-2 (current) and the 25 kGy dose is substantiated by VD<sub>25</sub> Method Max testing.

### Conformity to the European Directives

The devices have been classified as Class III, in accordance with Annex IX of the 93/42/EEC Medical Devices Directive., as they are manufactured utilising animal tissues or derivative which are rendered non-viable during the process – i.e. one of the minor (but crucial) transient processing aids is derived from an animal substance (Rule 17).