

## Suction Sets Range Specification Sheet



### Product Overview

The NETWORK ENT® Suction Sets range, has been specifically designed to provide precision and control during ENT procedures.

- Complete suction system includes suction tube with handle, tubing and adapter
- Compatible with all commercially available suction systems
- Supplied sterile, single use only, declared 5 year shelf life
- Manufactured in the UK

### Suction Sets Range Size Options

Product Size	Box Quantity	Product Code
Suction set complete with house adapter, with suction control hole	50	74-2003
Suction set with Zoellner type 2mm x 85mm, 10 degree shallow distal bend (2.0mm/6Fr)	50	74-2004
Suction set with Zoellner type 2mm x 85mm, 30 degree acute distal bend (2.0mm/6Fr)	50	74-2005
Suction set with Belucci type 2mm x 85mm, proximal bend, with suction control (2.0mm/6Fr)	50	74-2006

## Suction Sets Range Specification Sheet

### Material Specification

Product Component	Specification
Tube (74-2004,74-2005 & 74-2006 only)	Stainless Steel
Moulded handle	Polypropylene (PP)
House Adapter (74-2003 only)	ABS
Reducer connector	Polyamide
Connector Tubing	Silicone
Pouch Packaging	Tyvek/Film
Outer Box/Carton Packaging	500 Micron Brown Boxboard

### Intended Use

The NETWORK ENT® Suction Sets are designed specifically for ENT procedures and are intended for the removal of excess fluids and debris from the surgical site.

### Sterilisation

Products are sterilised by Ethylene Oxide (EO) according to a validated cycle and the requirements of BS EN ISO 11135-1(current).

### Instructions for Use

The instructions for use and suggested surgical technique are supplied in the form of a multi-lingual leaflet with instructions given in diagram form where appropriate. The international symbols used on the packaging are explained in each language. A leaflet is supplied with each product.

### Conformity to the European Directives

The Network range of Aspiration Suction Tubes and Fine Ends are identified as surgically invasive devices for transient use and are therefore classified as a Class IIa devices according to Annex IX, Rule 6, of the Medical Devices Directive 93/42/EEC as amended by 2007/47/EC.